

LOWENSTEIN JENSEN MEDIUM W/ ETHAMBUTOL 2 µG/ML

INTENDED USE:

Lowenstein-Jensen Medium is commonly used in the clinical laboratory for the isolation and differentiation of Mycobacterium spp.

PRINCIPLE AND INTERPRETATION:

L-Asparagine and Potato Flour are sources of nitrogen and vitamins in Lowenstein-Jensen Medium. Monopotassium Phosphate and Magnesium Sulfate enhance organism growth and act as buffers. Glycerol and the Egg Suspension provide fatty acids and protein required for the metabolism of mycobacteria. The coagulation of the egg albumin during sterilization provides a solid medium for inoculation purposes. Sodium Citrate and Malachite Green are selective agents to prevent growth of most contaminants and allow early growth of mycobacteria.

COMPOSITION:

| Ingredients | Gr/1600 ml |
|-------------------------|------------|
| L-Asparagine | 3,6 gr |
| Monopotassium Phosphate | 2,5 gr |
| Magnesium Sulfate | 0,24 gr |
| Sodium Citrate | 0,6 gr |
| Malachite Green | 0,4 gr |
| Potato Flour | 30 gr |
| Glycerol | 12 ml |
| Egg Suspension | 1000 ml |
| Ethambutol | 0,0032gr |
| Deionize water | 600 ml |

***Formula adjusted, standardized to suit performance parameters

pH: 7,0 ± 0,2

PRECAUTIONS:

For professional use only. Do not use tubes if they show evidence of microbial contamination, discoloration, or other signs of deterioration.

TEST PROCEDURE:

Material Provided: Lowenstein Jensen Medium w/ Ethambutol 2 µg/ml

| Materials Required but Not Provided | Quantity |
|---------------------------------------|----------|
| Suspension Tube | 1 |
| 10 ⁻² dilution tube (10ml) | 1 |
| 10 ⁻⁴ dilution tube (10ml) | 1 |

- 1.Suspend the sample in a sterile screw-capped glass tube containing sterile glass beads.
- 2.Vortex well (several minutes) until suspension is free of large clumps.
- 3.Compare this suspension to 1 McFarland. The suspension should be more turbid than the standard.
- 4.Take 100 µl from 1 McFarland suspension tube and pour in to 10⁻² dilution tube and vortex.
- 5.Take 100 µl from 10⁻² dilution tube and pour in to 10⁻⁴ dilution tube and vortex.
- 6.Take 100 µl from 10⁻⁴ dilution tube and inoculate in to Lowenstein Jensen Medium w/ Ethambutol 2 µg/ml
- 7.Incubate tubes with loosened caps at 35 ± 2 °C in an aerobic atmosphere supplemented with 5-10% carbon dioxide.
- 8.Examine tubes after 7, 14, and if necessary, 21 days for amount of growth and inhibition.
- 9.The growths are evaluated according to below table:
 - R (resistant); the bacteria are resistant and it are growing.
 - S (sensitive); the bacteria are sensitive and are not growing.

QUALITY CONTROL:

1.Sterility Control:

30-35°C / 21 d: NO GROWTH

2.Physical/Chemical Control

pH: 7,0 ± 0,2

Apperance: Blue-Green

3.Microbiological Control: Incubation at 35±2°C and examine tubes after 7, 14, and 21 days.

| Microorganism | Results | |
|----------------------------------|--------------------|--------------------|
| | Growth | Reaction |
| M. tuberculosis H37Ra ATCC 25177 | Inhibition | Sensitive |
| M. smegmatis ATCC 607 | Inhibition | Sensitive |
| Escherichia coli 25922 | Partial inhibition | Partial inhibition |

**Incubate uninoculated representative tubes aerobically at 20-25 °C and 30-35 °C and examine after 7 and 14 days for microbial contamination.

STORAGE CONDITIONS AND SHELF LIFE:

Store the prepared medium at 2-12°C. Use before expiry date on the label. Do not use beyond stated expiry date.

DISPOSAL:

Incubated medium may contain active bacteria and micro-organisms. Do not open infected medium. Infected tube should be autoclaved, incinerated or opened and soaked in a chlorine-based disinfectant (liquid bleach) for 20 minutes prior to disposal.

PACKAGING:

Katalog Number: 04015

Content/Packaging: 16x160mm Glass Tube, 100 tubes/box

REFERENCES:

- 1.Musser, J. M. 1995. Antimicrobial resistance in Mycobacteria: molecular genetic insights. Clinical Microbiology Reviews. 8:496-514.
2. Murray, P. R., E. J. Baron, M. A. Pfaller, F. C. Tenover, and R. H. Tenover (eds.). 1995. Manual of clinical microbiology, 6th ed. American Society for Microbiology, Washington, D.C.
3. Lowenstein, E. 1931. Die Zachtung der Tuberkelbazillen aus dem stramenden Blute. Zentralb. Bakteriol Parasitenkd. infektionskr. hyg. Abt. I orig., 120:127.
4. Jensen, K. A. 1932. Rinzuchtung und Typenbestimmung von Tuberkelbazillentamen. Zentralb. Bakteriol Parasitenkd. infektionskr. hyg. Agt. I Orig., 125:222.
5. MacFaddin, J. F. 1985. Media for isolation-cultivation-identification-maintenance of medical bacteria, vol. 1. Williams & Wilkins, Baltimore, MD.
6. Isenberg, H. D. (ed.). 1992. Clinical microbiology procedures handbook, vol. 1 American Society for Microbiology, Washington, D.C.



Aseptic Sterile



Batch Code



Catalogue Number



Negative Controls



Positive Controls



Use by



Temperature
Limitation



Do not reuse



Contains sufficient
for <n> tests



Look at user manual



Manufacturer