



## CERTIFICATE OF ANALYSIS

<b>Product Name</b>	Phosphate Buffered Peptone Saline(PBPS) + %3 Tween80 + %0,3 Lecithin (90mL)
<b>Catalogue No</b>	01152
<b>Batch No</b>	1905159
<b>Dehydrated Brand / Batch No / Expiry Date</b>	Buffered Saline Pepton Water (CONDALAB): 810104 / 10-2022
<b>Neutralizer Brand / Batch No / Expiry Date</b>	Tween 80 (SABO): 20103868 / 05-2020 Lecithin (LAB-M): Q42844 / 10-2020
<b>Production Date</b>	10.05.2019
<b>Expiry Date</b>	10.09.2019
<b>Storage Conditions</b>	2 – 25 °C

### Physical & Chemical Specifications

<b>Appearance</b>	clear	Appropriate
<b>Color</b>	Light amber	Appropriate
<b>pH</b>	7,0±0,2	7,06

### Microbiological Specifications: Productivity and acc. To ISO/TS 11133

<b>Method of control:</b> Direct Inoculation	<b>Inoculum:</b> :(+ control 10-100 cfu, - control 100-1000 cfu)	<b>Incubation:</b> 35±2 °C:18-24h			
<b>Sterility:</b>	<b>Incubation:</b> (20-25°C : 3 d, 30-35°C : 2 d)	Sterile			
<b>QC Strains</b>	<b>ATCC</b>	<b>Results</b>			
		<b>Recovery</b>		<b>Growth</b>	
<i>Staphylococcus aureus</i>	25923	Good	Good	Good	Appropriate
<i>Pseudomonas aeruginosa</i>	9027	Good	Good	Good	Appropriate
<i>Bacillus subtilis</i>	6633	Good	Good	Good	Appropriate
<i>Escherichia coli</i>	25922	Good	Good	Good	Appropriate
<i>Salmonella typhimurium</i>	14028	Good	Good	Good	Appropriate
<i>Candida albicans</i>	10231	Good	Good	Good	Appropriate

<b>QC Date</b>	<b>15.05.2019</b>
<b>QC Manager</b>	<b>İlknur ÇETİN</b>

This document has been prepared electronically.

RTA LABORATUVARLARI BİYOLOJİK ÜRÜNLER İLAÇ VE MAKİNE SANAYİ TİCARET A.Ş.  
Gebze Plastikçiler Organize Sanayi Bölgesi (GEPOSB) Cumhuriyet Cad.No.:3 41400 Gebze / KOCAELİ  
Tel.:0262 648 53 00 Fax:0262 751 06 77 e-mail: [rtalabs.com.tr](mailto:rtalabs.com.tr)